


FK-26888-000

THE FK-26888-000 IS AN UNDAMPED MAGNETIC BALANCED ARMATURE RECEIVER INTENDED FOR USE IN CIC HEARING INSTRUMENTS. THIS MODEL HAS HIGH PERMEABILITY CUP & COVER TO REDUCE MAGNETIC RADIATION, AND HAS THE CENTER TERMINAL ELECTRICALLY CONNECTED TO CASE.

CONSTANT VOLTAGE DRIVE CONDITIONS



## ACOUSTICAL

SENSITIVITY

DEVICE WILL PRODUCE THE SPL LISTED BELOW UNDER TEST CONDITIONS DESCRIBED IN TABLE 3. NOMINAL SENSITIVITY AT IKHZ IS dB RELATIVE TO 20µPa. ALL OTHER VALUES IN dB RELATIVE TO THE SENSITIVITY AT IKHZ.

| FREQUENCY (Hz) | MINIMUM | NOMINAL | MAXIMUM |
|----------------|---------|---------|---------|
| 200            | 0.0     | +2.0    | +4.0    |
| 500            | -0.5    | +1.5    | +3.5    |
| 1000           | -2.0    | 93.0    | +2.0    |
| 2250 - 3250    | +5.0    | +8.0    | +  .0   |
| 6000           | 80.0    |         |         |

TABLE I

TOTAL HARMONIC DISTORTION

DEVICE WILL NOT EXCEED TOTAL HARMONIC DISTORTION LEVELS LISTED BELOW.

| FREQUENCY (Hz) | AC DRIVE (V rms) | DC BIAS (mA) | LIMIT (%) |
|----------------|------------------|--------------|-----------|
| 920            | 0.367            | 0            | 5         |
| 1380           | 0.367            | 0            | 5         |
| 920            | 1.04             | 0            | 10        |

TABLE 2

## TEST CONDITIONS

| TEOT COMBITTORS        |   |
|------------------------|---|
| NOMINAL SOURCE VOLTAGE | 0.367 V rms, 0 mA DC BIAS                                       |
| SOURCE IMPEDANCE       | <  Ohm  |
| TUBING                 |   |
| COUPLER CAVITY         | 2 CM <sup>3</sup> , SIMULATED ANSI S3.7 TYPE HA-3 (IEC 60318-5) |

TABLE 3

## ELECTRICAL

| DC RESISTANCE      | 580 Ohms ± 10%     |
|--------------------|--------------------|
| IMPEDANCE @ 500 Hz | 690 Ohms $\pm$ 15% |
| IMPEDANCE @ IkHz   | 940 Ohms ± 15%     |

TABLE 4

ISOLATION: CASE WILL BE ELECTRICALLY ISOLATED FROM THE COIL CIRCUIT.

|          | E<br>D   | C10114032<br>C10112265 |                   |        | Active              |               | E                   |
|----------|--|------------------------|-------------------|--------|---------------------|---------------|---------------------|
| S        | WHEN TEST LIMITS ARE USED TO ESTABLISH INCOMING INSPECTION ACCEPTANCE/REJECTION CRITERIA, CORRELATION OF TEST EQUIPMENT WITH KNOWLES IS ALSO REQUIRED FOR ELIMINATION OF EQUIPMENT AND TEST METHOD VARIATION |                        |                   | DR. BY | DATE<br>7 - 31 - 06 |               |                     |
| <b>3</b> | TITLE:   | DE                     | CFIVFR            |        | K-26888-000         | CK. BY<br>GJP | DATE<br>8 - 4 - 0 6 |
|          |  | ΠL                     | CEIVER            | 「      | K-20000-000         | APP. BY       | DATE                |
|          | I  | PERFORMAN              | NCE SPECIFICATION |        | SHT 2 I             | GID           | 8-1-06              |

Revision C.O. # Implementation Date

PORT LOCATION: 3S

TEMPERATURE

SOLDER TYPE: SAC305

OPERATURE: SENSITIVITY WILL NOT VARY

-17°C TO 63°C.

STORAGE: -40°C TO 63°C.

MORE THAN +1/-3 dB FROM

| KNOWLES ELECTRONICS     |
|-------------------------|
| ITASCA, ILLINOIS U.S.A. |